



www.ICLabs.ca

Memo No. 2018-012
Date: 1-Mar-2018
Memo To: Clients
Re: **Assay Delay: Allergen C5 Ampicilloyl and Allergen C6 Amoxicilloyl**

Somagen Diagnostics has issued an urgent recall for Allergen C5 Ampicilloyl and Allergen C6 Amoxicilloyl reagents. The affected test results have been amended due to the reagent stability issue identified by the manufacturer of the diagnostic assay.

The recall has resulted in the reagent being placed on indefinite backorder and the tests will not be available until further notice.

If you have further questions please contact Client Care at (416) 422-3000 Ext. 300 or info@ICLabs.ca

Shashank Tilak.
Laboratory Director
(416) 422-3000 Ext. 221
ShashTilak@ICLabs.ca

Want to receive updates by e-mail? Please contact Client Care info@ICLabs.ca

ICL is the only Not-For-Profit dedicated broker of lab diagnostic tests in Canada
57 Gervais Drive, North York, ON M3C 1Z2 (416) 422-3000 Ext. 300 info@ICLabs.ca

Page 1 of 1